

1<sup>ST</sup> QUARTER 2006  
AIR QUALITY AND METEOROLOGICAL  
MONITORING AUDIT REPORT

PREPARED FOR:

JIM SICKLES  
US ENVIRONMENTAL PROTECTION AGENCY  
REGION 9  
REMEDIAL PROJECT MANAGER



YERINGTON/ANACONDA MINE SITE  
YERINGTON, NEVADA

*Prepared by*



**Tetra Tech EM Inc.**  
135 Main St., Suite 1800  
San Francisco, California 94105

**May 8, 2006**

## TABLE OF CONTENTS

<u>Section</u>	<u>Page</u>
1.0 Oversight and Audit Summary .....	1
2.0 PM <sub>10</sub> , TSP, and Meteorological Audit Methods .....	3
3.0 Meteorological Audit Equipment .....	9
4.0 Meteorological and Air Quality Audit Summary of Results and Recommendations .....	10

## LIST OF TABLES

<u>Table</u>	<u>Page</u>
1-1 Air Quality and Meteorological Monitoring Parameters Yerington Mine Site .....	1
2-1 Summary of PM <sub>10</sub> Audit Results Yerington Mine Site January 9 -10, 2006 .....	4
2-2 Summary of TSP Audit Results Yerington Mine Site January 9 -10, 2006 .....	5
2-3 PSD Quality Assurance Audit Methods, Accuracy Requirements, and Results Yerington Meteorological Tower Audit January 10, 2006.....	8
3-1 Quality Assurance Audit Equipment Yerington PM <sub>10</sub> , TSP, and Meteorological Tower Audits January 9 -10, 2006	

## LIST OF APPENDICES

Appendix A	Quality Assurance Audit Data Tables
Appendix B	Audit Equipment Standards Certifications and Field Logbook Notes

## 1.0 Oversight and Audit Summary

On behalf of the US Environmental Protection Agency (EPA), Tetra Tech EM Inc. (Tetra Tech) personnel are providing ongoing regulatory support for the Yerington Mine air quality and meteorological monitoring program. Atlantic Richfield Company's (ARC) environmental contractor, Brown and Caldwell (B & C), are currently performing all aspects of this program. Tetra Tech reviewed and provided comments and feedback on the Air Quality Monitoring Work Plan for the Yerington Mine Site (AQMWP), Authored by B & C.

The AQMWP provides details on proposed actions for air quality and meteorological monitoring at the Yerington site. The air quality-monitoring component of the AQMWP includes the installation and operation of seven Hi-Volume (Hi-Vol) particulate matter less than 10 microns (PM<sub>10</sub>) samplers and six Hi-Vol total suspended particulate (TSP) samplers at sample locations approved by ARC, B&C, EPA, and Tetra Tech. The meteorological monitoring component of the AQMWP includes the continued operation of an existing 10-foot meteorological tower. The air quality and meteorological monitoring parameters and instrumentation for the Yerington Mine Site are summarized in Table 1-1.

**Table 1-1**

### **Air Quality and Meteorological Monitoring Parameters for Yerington Mine Site**

<b>Parameter</b>	<b>Instrument Description-Model</b>	<b>Instrument Location</b>
Wind Speed	RM Young Model 05305 AQ	10 feet AGL
Wind Direction	RM Young Model 05305 AQ	10 feet AGL
Ambient Temperature	Vaisala Temperature/Humidity Probe	7 feet AGL
Relative Humidity	Vaisala Temperature/Humidity Probe	7 feet AGL
Solar Radiation	Licor Model LI200X	9 feet AGL
Station Barometric Pressure Sensor	Vaisala Barometric Pressure Sensor	Meteorological Tower Enclosure Cabinet
Meteorological Tower Datalogger	Campbell Scientific CR10X	Meteorological Tower Enclosure Cabinet
PM <sub>10</sub> Hi-Volume FRM Sampler	Tisch Environmental, Inc. Model TE-6070D	7 Total samplers at 6 locations
TSP Hi-Volume FRM Sampler	Tisch Environmental, Inc. Model TE-5170D	6 Total samplers at 6 locations

## Table 1-1 (continued)

### Notes:

AGL	Above ground level
FRM	Federal Reference Method
PM <sub>10</sub>	Particulate matter less than 10 microns in diameter
TSP	Total suspended particulate

Tetra Tech personnel coordinated with B&C personnel to meet at the site and perform EPA-reference method audit procedures on seven PM<sub>10</sub> samplers, six TSP samplers, and the on-site meteorological tower. PM<sub>10</sub> and TSP audits were performed using a certified Hi-flow audit orifice and the meteorological tower was audited using certified reference sensors.

Mr. Doug Herlocker of Tetra Tech conducted the audit procedures at the Yerington site on January 9 through 10, 2006. The following tasks were performed:

- EPA and manufacturer-approved audit of seven PM<sub>10</sub> Hi-Vol and six TSP Hi-Vol samplers using certified audit orifice (Completed January 9 through 10, 2006)
- Prevention of Significant Deterioration (PSD)-quality audit of 10-foot meteorological tower (Completed January 10, 2006)

All PM<sub>10</sub> and TSP samplers audited on January 9 through 10, 2006, successfully passed all audit criteria. No equipment failures, leaks, or anomalies were observed during the audit procedure.

At the time of the meteorological tower audit on January 10, 2006, all sensors were operating within PSD-audit criteria. During the previous audit, a damaged precipitation gauge wire was identified. The damaged wire was observed to be replaced with a new wire and was operating correctly at the time of the audit.

Details of the quality oversight and audit summary are presented in the following sections and appendices:

- Section 2.0 PM<sub>10</sub>, TSP, and Meteorological Audit Methods
- Section 3.0 Meteorological Audit Equipment
- Section 4.0 Meteorological and Air Quality Audit Summary of Results and Recommendations
- Appendix A Quality Assurance Audit Data Tables
- Appendix B Audit Equipment Standards Certifications and Field Logbook Notes

## 2.0 PM<sub>10</sub>, TSP, and Meteorological Audit Methods

Tetra Tech personnel audited all seven PM<sub>10</sub> samplers, all six TSP samplers, and the 10-foot meteorological tower on January 9-10, 2006.

The Yerington 10-foot meteorological tower sensors were audited in their normal operating modes. The accuracy of all sensors was verified using National Institute of Standards and Technology (NIST) or Certified Reference Material (CRM) traceable transfer standard reference audit sensors.

A description of audit procedures and methods is presented below.

### **PM<sub>10</sub> Hi-Vol Audit Procedure**

Seven PM<sub>10</sub> Hi-Vol samplers are installed at six locations at or near the perimeter of the Yerington mine site. The samplers operate by activation of a pump via a timer system. The audit device consists of a certified audit orifice (orifice) designed for PM<sub>10</sub>/TSP samplers. The orifice consists of metal plate attached to a flow restriction cylinder, which allows the user to turn a knob and restrict flow to the sampler. In addition, the orifice is equipped with a manometer port to measure pressure drop through the orifice.

The orifice is designed to produce flow ranges of 1.02 to 1.24 cubic meters per minute (m<sup>3</sup>/min) based on the design flow rate of the PM<sub>10</sub> Hi-Vol samplers. Five different flow measurements are obtained using the orifice and resultant manometer readings are recorded. PM<sub>10</sub> sampler flow readings are also recorded for the five different flow rates. In addition, ambient temperature and pressure are recorded. These site-specific parameters were obtained during the audit using NIST certified devices. A slope, intercept, and correlation coefficient is calculated using the least squares regression method and is based on the five orifice manometer and sampler flow readings. The audit criteria are based on a correlation coefficient of 0.990 or higher. A summary of PM<sub>10</sub> audit results is presented in Table 2-1. Quality assurance audit data tables are presented in Appendix A.

**TABLE 2-1**

**SUMMARY OF PM<sub>10</sub> AUDIT RESULTS  
YERINGTON MINE SITE  
JANUARY 9 -10, 2006**

<b>PM<sub>10</sub> Sampler/ Location</b>	<b>Sampler Serial Number</b>	<b>Sampler Orifice Serial Number</b>	<b>Audit Orifice Serial Number</b>	<b>Ambient Temperature (degrees C<sup>o</sup>)</b>	<b>Ambient Pressure (mm Hg)</b>	<b>Sampler/Audit Orifice Correlation Coefficient</b>
AM1	613	1013	W43	13.7	649.7	0.9967
AM1-DUP	616	1022	W43	12.9	649.7	0.9947
AM2	617	1016	W43	9.5	649.7	0.9966
AM3	618	1018	W43	9.1	653.0	0.9901
AM4	619	1012	W43	3.1	653.0	0.9943
AM5	614	1019	W43	7.7	653.0	0.9993
AM6	615	1020	W43	9.8	654.8	0.9930

Notes:

- °C                    Degree Celsius
- DUP                 Duplicate
- mm Hg              Millimeter mercury
- PM<sub>10</sub>                Particulate matter less than 10 microns in diameter

**TSP Hi-Vol Audit Procedure**

Six TSP Hi-Vol samplers are collocated with the PM<sub>10</sub> samplers and installed at six locations at or near the perimeter of the Yerington mine site. The samplers operate by the activation of a pump via a timer system, identical to the PM<sub>10</sub> samplers. The audit device consists of a certified audit orifice (orifice) designed for PM<sub>10</sub>/TSP samplers. The orifice consists of metal plate attached to a flow restriction cylinder, which allows the user to turn a knob and restrict flow to the sampler. In addition, the orifice is equipped with a manometer port to measure pressure drop through the orifice.

The orifice is designed to produce flow ranges of 1.10 to 1.70 cubic meters per minute (m<sup>3</sup>/min) based on the design flow rate of the TSP Hi-Vol samplers. Five different flow measurements are obtained using the orifice and resultant manometer readings are recorded. TSP sampler flow readings are also recorded for the five different flow rates. In addition, ambient temperature and pressure are recorded. These site-specific parameters were obtained during the audit using NIST certified devices. A slope, intercept, and correlation coefficient is calculated using the least squares regression method and is based on the five orifice manometer and sampler flow readings. The audit criteria are based on a correlation coefficient of 0.990 or higher. A summary of TSP audit results is presented in Table 2-2. Quality assurance audit data tables are presented in Appendix A.

**TABLE 2-2**  
**SUMMARY OF TSP AUDIT RESULTS**  
**YERINGTON MINE SITE**  
**JANUARY 9 -10, 2006**

<b>TSP Sampler/ Location</b>	<b>Sampler Serial Number</b>	<b>Sampler Orifice Serial Number</b>	<b>Audit Orifice Serial Number</b>	<b>Ambient Temperature (degrees C<sup>o</sup>)</b>	<b>Ambient Pressure (mm Hg)</b>	<b>Sampler/Audit Orifice Correlation Coefficient</b>
AM1	NA	1033	W43	9.8	649.7	0.9992
AM2	NA	1034	W43	9.5	649.7	0.9941
AM3	NA	1014	W43	9.4	653.0	0.9995
AM4	NA	1015	W43	3.1	653.0	0.9990
AM5	NA	1017	W43	7.7	653.0	0.9962
AM6	NA	1021	W43	9.8	654.8	0.9979

Notes:

- C                    Degree Celsius
- NA                   Not available
- mm Hg              Millimeter mercury
- TSP                  Total suspended particulate

**Vane Alignment Verification/Audit**

The Yerington meteorological tower wind speed/direction sensor is mounted at the top of the tower (10-foot level) on a 1-meter cross arm fixture oriented in a north/south direction, equivalent to a value of zero degrees. The vane alignment audit was achieved using a calibrated compass set on a tripod and corrected for true north offset for the Yerington, Nevada area of approximately 15.3 degrees Easterly. An additional audit was achieved using a handheld global positioning system (GPS), which automatically corrects for true north readings. Audit devices were aligned with the cross arm facing north. The orientation of the cross arm was then compared to the audit device reading. The results from the true north vane alignment audit are presented below:

- Audit compass (corrected 15.3 degrees Easterly) = 359 degrees
- Audit GPS (automatically corrected for true north) = 0.6 degrees
- Wind direction sensor aligned north-facing = 1.8 degrees

### **Wind Speed Audit**

The wind speed audit was achieved using a R.M. Young Motor Drive (Model No. 18802) attached to the wind speed/direction sensor and was rotated at different speeds. The simulated wind speed was compared to datalogger readouts for wind speed accuracy.

In addition, a wind speed starting threshold torque audit was performed using a R.M. Young Torque Disc (Model No. 18310) to verify the sensitivity of the wind speed sensor to fluctuations in wind speed. Acceptance and accuracy criteria are presented in Table 2-3.

### **Wind Direction Audit**

The wind direction audit was achieved using a R.M. Young Wind Direction Linearity Gauge (Model No. 18802) attached to the wind speed/direction sensor and was rotated between 0 and 360 degrees in 30-degree increments. The circular gauge has marks in 1-degree increments that were compared to the datalogger readouts for wind direction accuracy.

In addition, a wind direction starting torque audit was performed using a R.M. Young Wind Direction Torque Gauge (Model No. 18331) to verify the sensitivity of the wind direction sensor to fluctuations in wind speed. Acceptance and accuracy criteria are presented in Table 2-3.

### **Temperature Audit**

The temperature audit was achieved using a NIST traceable reference temperature sensor collocated with the 7-foot temperature sensor. Ambient temperature readings were recorded for both sensors. The NIST traceable reference temperature sensor was compared to the 7-foot sensor for temperature accuracy. Acceptance and accuracy criteria are presented in Table 2-3.

### **Humidity Audit**

The humidity audit was achieved using a NIST traceable reference humidity sensor (hygrometer) collocated with tower humidity sensor. Ambient humidity conditions were recorded for both sensors and dewpoint temperatures were calculated for the both sensors using ambient temperature and humidity readouts, and were compared for humidity accuracy. Acceptance and accuracy criteria are presented in Table 2-3.

### **Solar Radiation Audit**

The solar radiation audit was achieved using a CRM traceable reference solar radiation sensor collocated with tower sensor. Four short-term solar radiation readouts were recorded for both sensors and were compared for solar radiation accuracy. Acceptance and accuracy criteria are presented in Table 2-3.

**Precipitation Audit**

The precipitation audit was achieved using a precise liquid dispensing tool to measure the amount of water required to initiate a tip for each tipping bucket mechanism. Three filling runs were performed for each bucket and compared to specified volume required for one tip. Acceptance and accuracy criteria are presented in Table 2-3.

**Barometric Pressure Audit**

The barometric pressure audit was achieved using a NIST traceable barometric pressure sensor collocated with tower sensor. Three readouts were recorded for both sensors and were compared for barometric pressure accuracy. Acceptance and accuracy criteria are presented in Table 2-3.

All quality assurance audit methods, accuracy requirements, and audit (pass/fail) results for the Yerington meteorological tower are summarized in Table 2-3. Quality assurance audit data tables are presented in Appendix A. Field logbook notes are presented in Appendix B.

**Table 2-3**

**PSD Quality Assurance Audit Methods, Accuracy Requirements, and Results  
Yerington Meteorological Tower Audit January 10, 2006**

<b>Parameter</b>	<b>Audit Method</b>	<b>Accuracy Requirements</b> (difference between acceptable criteria and sensor response)	<b>Within Acceptance Criteria</b>
10-Foot Wind Speed	Active Rotation with Certified Drive Unit: ws $\leq$ 5m/s ws $>$ 5m/s	$\leq \pm 0.25$ m/s $\leq \pm 5.0\%$	Yes
	Starting Threshold with Torque Disk	$\leq 0.5$ m/s (0.3 gm-cm)	Yes
10-Foot Wind Direction	Alignment Verification	$\leq \pm 5^\circ$ of True North	Yes
	Internal Check of Vane Linearity using Gauge	$\leq \pm 3^\circ$	Yes
	Starting Threshold with Torque Gauge	$\leq 0.5$ m/s (9 gm-cm)	Yes
7-Foot Ambient Temperature	Collocated Sensor Comparing Temperatures Using Three Water Baths	$\leq \pm 0.5$ °C	Yes
7-Foot Relative Humidity	Collocated Sensor Comparing Dewpoint	$\leq \pm 1.5$ °C Error in	Yes

	Temperatures (T <sub>dp</sub> )	T <sub>dp</sub>	
2-Meter Solar Radiation	Collocated Sensor Comparison	$\leq \pm 5.0\%$ Full Scale	Yes
Barometric Pressure	Collocated Sensor Comparison	$\leq \pm 10$ millibars Hg	Yes
Precipitation	Comparison to Precipitation Gauge	$\leq \pm 10.0\%$	Yes

Notes:

°	degree
°C	degree Celsius
gm-cm	gram-centimeter
m/s	meter per second
PSD	Prevention of significant deterioration
T <sub>dp</sub>	dewpoint temperature
ws	wind speed

### 3.0 Meteorological Audit Equipment

All audit equipment and reference standards were in current calibration and traceable to the NIST or other authoritative references. Table 3-1 lists specific equipment used and certification dates. Copies of standard certifications for the audit equipment are presented in Appendix B.

**Table 3-1**

**Quality Assurance Audit Equipment  
Yerington PM<sub>10</sub>, TSP, and Meteorological Tower Audits January 9 - 10, 2006**

References/Device	Manufacturer	Model Number	Serial Number	Recertification Date
PM <sub>10</sub> /TSP Audit Orifice	Tisch Environmental	TE-5028A	W43	4/06/3006
Humidity	Control Company	11-661-18	41531319	1/28/2007
Thermometer	Control Company	11-661-18	41531319	1/28/2007
Wind Speed Drive	RM Young	18802	CA02612	4/13/2006
Wind Direction Linearity Gauge	RM Young	18212	N/A	N/A
Wind Speed Starting Threshold Torque Disk	RM Young	18310	N/A	N/A
Wind Direction Starting Threshold Torque Gauge	RM Young	18331	N/A	N/A
Solar Radiation	Li-Cor	LI-200SZ	PY47392	8/13/2007
Barometric Pressure	Brunton	Multi-Navigator V 2.16	ACP 010796	4/13/2006

Notes:

N/A            Not available  
 PM<sub>10</sub>        Particulate matter less than 10 microns in diameter  
 TSP           Total suspended particulate

## **4.0 Meteorological and Air Quality Audit Summary of Results and Recommendations**

### PM<sub>10</sub> and TSP Audit Summary:

At the time of the PM<sub>10</sub> and TSP audit on January 9 through 10, 2006, all samplers successfully passed audit parameters and were observed to be operating correctly.

### Meteorological Tower Audit:

At the time of the meteorological tower audit on January 10, 2006, all NIST and CRM reference sensors and meteorological tower sensors were compared to the accuracy requirements established in Table 2-3 and all sensors were observed to be operating within accuracy requirements.

### General Recommendations:

Tetra Tech recommends that B&C continue to perform scheduled bi-weekly site visits to the meteorological tower to download data and visually inspect the sensors. In addition, downloaded data should be screened and evaluated within 48 hours to identify problems and minimize loss of data.

**APPENDIX A**  
**QUALITY ASSURANCE AUDIT DATA TABLES**

TABLE 1 WIND SPEED/DIRECTION RESPONSE AUDIT						
YERINGTON METEOROLOGICAL TOWER AUDIT AUDIT DATE: January 10, 2006 SITE: Yerington Mine Site, Yerington, NV AUDITED BY: Doug Herlocker, Tetra Tech EM Inc. on behalf of U.S. EPA						
WIND SPEED: (MODEL: RM Young 05305 AQ)				WIND DIRECTION: (MODEL: RM Young 05305 AQ)		
Audit Device: RM Young Model 18802 (Serial No. CA02415)				Audit Device: RM Young Model 18212 (Serial No. N/A)		
Acceptable Difference: ws ≤ 11.0 mph = 0.56 mph, ws > 11.0 mph = ≤ 5.0%				Acceptable Difference: ± 3 degrees		
Clockwise Rotation				Clockwise Rotation		
Calibration Device RPM	Simulated Wind Speed (mph)	ws Sensor <sup>a</sup> as found (mph)	Difference	Calibration Device (degrees)	wd Sensor as found (degrees)	Difference (degrees)
0.0	0.00	0.00	0.00 mph	0.0	1.8	1.8
200.0	2.19	2.28	0.09 mph	30.0	31.8	1.8
400.0	4.38	4.56	0.18 mph	60.0	61.5	1.5
600.0	6.58	6.84	0.26 mph	90.0	90.8	0.8
800.0	8.77	9.12	0.35 mph	120.0	120.3	0.3
1000.0	10.96	11.40	0.44 mph	150.0	150.2	0.2
1600.0	17.54	18.24	4.0 %	180.0	179.7	-0.3
2200.0	24.11	25.08	4.0 %	210.0	210.0	0.0
2600.0	28.50	29.64	4.0 %	240.0	239.9	-0.1
3000.0	32.88	34.20	4.0 %	270.0	271.6	1.6
Criteria met (yes/no): YES				300.0	301.2	1.2
Adjustment performed (yes/no): YES				330.0	330.9	0.9
				360/0.0	2.40	2.4
				Criteria met (yes/no): YES		
				Adjustment performed (yes/no): YES		

Notes:

<sup>a</sup> RM Young wind speed multiplier used with calibration device: (RPM x 0.01096)=mph, (RPM x 0.005)=m/s

<sup>b</sup> Wind Speed less than or equal to 11 miles per hour is compared to actual simulated wind speed miles per hour.

<sup>c</sup> Wind Speed greater than 11 miles per hour is compared to percent difference of simulated wind speed miles per hour.

% Percent

EPA U.S. Environmental Protection Agency

mph Mile per hour

m/s meter per second

N/A Not available

RPM Revolutions per minute

wd Wind direction

ws Wind speed

TABLE 2 WIND DIRECTION/SPEED STARTING THRESHOLD TORQUE AUDIT					
YERINGTON METEOROLOGICAL TOWER AUDIT AUDIT DATE: January 10, 2006 SITE: Yerington Mine Site, Yerington, NV AUDITED BY: Doug Herlocker, Tetra Tech EM Inc. on behalf of U.S. EPA					
WIND SPEED STARTING TORQUE THRESHOLD: (MODEL: RM Young 05305 AQ)			WIND DIRECTION STARTING TORQUE THRESHOLD: (MODEL: RM Young 05305 AQ)		
Audit Device: RM Young Model 18310 (Serial No. N/A)			Audit Device: RM Young Model 18331 (Serial No. N/A)		
Acceptable Difference: ≤ 0.3 g-cm			Acceptable Difference: ≤ 9 g-cm		
Sensor	ws Sensor as found (g-cm)	Sensor	wd Sensor as found (g-cm)	Sensor	wd Sensor as found (g-cm)
10 meter ws	0.1	10 meter wd	N/A		
Sensor	ws Sensor as left (g-cm)	Sensor	wd Sensor as left (g-cm)	Sensor	wd Sensor as left (g-cm)
10 meter ws	0.1	10 meter wd	N/A		
Criteria met (yes/no): NO			Criteria met (yes/no): NO		
Adjustment performed (yes/no): NO			Adjustment performed (yes/no): NO		

Notes:

EPA U.S. Environmental Protection Agency

g-cm Gram-centimeter

N/A Not available

ws Wind speed

wd Wind direction

**TABLE 3  
HUMIDITY AUDIT**

YERINGTON METEOROLOGICAL TOWER AUDIT					
AUDIT DATE: <u>January 10, 2006</u>					
SITE: <u>Yerington Mine Site, Yerington, NV</u>					
AUDITED BY: <u>Doug Herlocker, Tetra Tech EM Inc. on behalf of U.S. EPA</u>					
HUMIDITY: Vaisala Model HMP45C with Campbell CR10X Datalogger					
Audit Device <sup>a</sup> : Control Company Model No. 11-661-18 (Serial No. 41531319)					
Acceptable Difference: +/- 4.1 °F in Dewpoint Temperature					
Audit device (% humidity)	Humidity Sensor as found (% humidity)	Difference (% humidity)	Audit device Dewpoint Temperature (°F)	Humidity Sensor Dewpoint Temperature (°F)	Dewpoint Difference Temperature (°F)
40.1	40.4	0.3	35.40	35.10	-0.30
Criteria met (yes/no): YES			Adjustment performed (yes/no): NO		

Notes:  
<sup>a</sup> Audit device meets National Institute of Standards and Technology requirements.  
 % Percent  
 °F Fahrenheit  
 EPA U.S. Environmental Protection Agency

**TABLE 4  
AMBIENT TEMPERATURE AUDIT**

YERINGTON METEOROLOGICAL TOWER AUDIT			
AUDIT DATE: <u>January 10, 2006</u>			
SITE: <u>Yerington Mine Site, Yerington, NV</u>			
AUDITED BY: <u>Doug Herlocker, Tetra Tech EM Inc. on behalf of U.S. EPA</u>			
TEMPERATURE: Vaisala Model HMP45C with Campbell CR10X Datalogger			
Audit Device <sup>a</sup> : Control Company Model No. 11-661-18 (Serial No. 41531319)			
Acceptable Difference: +/- 1.0 °F Mean Error			
Temperature Range	Audit Device Temperature (°F)	Ambient Temperature as found (°F)	Difference (°F)
Ambient range	53.1	52.6	-0.5
Mean Error =			-0.5
Criteria met (yes/no): YES		Adjustment performed (yes/no): NO	

Notes:  
<sup>a</sup> Audit device meets National Institute of Standards and Technology requirements.  
 °F Fahrenheit  
 EPA U.S. Environmental Protection Agency

**TABLE 5  
SOLAR RADIATION AUDIT**

YERINGTON METEOROLOGICAL TOWER AUDIT					
AUDIT DATE: <u>January 10, 2006</u>					
SITE: <u>Yerington Mine Site, Yerington, NV</u>					
AUDITED BY: <u>Doug Herlocker, Tetra Tech EM Inc. on behalf of U.S. EPA</u>					
SOLAR RADIATION: Kipp & Zonen with Campbell CR10X Datalogger					
Audit Device <sup>a</sup> : Licor Model LI200SZ (serial #PY47392)					
Acceptable Difference: < 5% (Full Scale)					
Audit Device Zero <sup>b</sup> (W/m <sup>2</sup> )	Sensor Device Zero <sup>b</sup> (W/m <sup>2</sup> )	Audit Device 1st. Reading (W/m <sup>2</sup> )	Sensor Device 1st. Reading (W/m <sup>2</sup> )	Audit Device 2nd. Reading (W/m <sup>2</sup> )	Sensor Device 2nd. Reading (W/m <sup>2</sup> )
0.0	0.0	419.0	434.0	418.0	432.0
Criteria met (yes/no): YES			Adjustment performed (yes/no): NO		

Notes:  
<sup>a</sup> Audit device meets National Institute of Standards and Technology requirements.  
<sup>b</sup> Audit and sensor devices covered to simulate night time environment.  
 EPA U.S. Environmental Protection Agency  
 W/m<sup>2</sup> Watts per meter squared  
 N/A Not available

**TABLE 6  
PRECIPITATION AUDIT**

**YERINGTON METEOROLOGICAL TOWER AUDIT**

**AUDIT DATE:** January 10, 2006

**SITE:** Yerington Mine Site, Yerington, NV

**AUDITED BY:** Doug Herlocker, Tetra Tech EM Inc. on behalf of U.S. EPA

**PRECIPITATION: Climatronics Model 100508 (Serial No. 935) with  
Campbell CR510 Datalogger (Serial No. 18)**

**Audit device: 10 mL syringe (Serial No. N/A )**

**Acceptable Difference: +/- 10% (1 tip = 8.3 mL = 0.01 inches of precipitation),  
<8.0 mL, > 8.6 mL = adjustment recommended**

<b>Volume Checks (number)</b>	<b>Left Bucket <i>as found</i></b>	<b>Right Bucket <i>as found</i></b>
1	8.6	8.0
2	8.4	8.0
3	8.3	8.1
<b>Average =</b>	<b>8.4</b>	<b>8.0</b>
<b>Volume Checks (number)</b>	<b>Left Bucket <i>as left</i></b>	<b>Right Bucket <i>as left</i></b>
1	8.6	8.0
2	8.4	8.0
3	8.3	8.1
<b>Average =</b>	<b>8.4</b>	<b>8.0</b>
	<b>Criteria met (yes/no): YES</b>	<b>Criteria met (yes/no): YES</b>
	<b>Adjustment performed (yes/no): NO</b>	<b>Adjustment performed (yes/no): NO</b>

**Notes:**

- %                      Percent
- EPA                    U.S. Environmental Protection Agency
- mL                     Milliliter
- N/A                    Not available

**APPENDIX B**

**AUDIT EQUIPMENT STANDARDS CERTIFICATIONS AND FIELD  
LOGBOOK NOTES**



TISCH ENVIRONMENTAL, INC.  
 145 SOUTH MIAMI AVE.  
 VILLAGE OF CLEVELAND, OH 45002  
 513.467.9000  
 877.253.7610 TOLL FREE  
 513.467.9009 FAX  
 WWW.TISCH-ENV.COM

AIR POLLUTION MONITORING EQUIPMENT

ORIFICE TRANSFER STANDARD CERTIFICATION WORKSHEET TE-5028A

Date - Apr 06, 2005 Rootsmeter S/N 9833620 Ta (K) - 295  
 Operator Tisch Orifice I.D. - W43 Pa (mm) - 753.11

PLATE OR VDC #	VOLUME START (m3)	VOLUME STOP (m3)	DIFF VOLUME (m3)	DIFF TIME (min)	METER DIFF Hg (mm)	ORFICE DIFF H2O (in.)
1	NA	NA	1.00	1.2560	4.5	1.50
2	NA	NA	1.00	0.9750	7.4	2.50
3	NA	NA	1.00	0.8880	8.9	3.00
4	NA	NA	1.00	0.8200	10.3	3.50
5	NA	NA	1.00	0.6190	17.7	6.00

DATA TABULATION

Vstd	(x axis) Qstd	(y axis)	Va	(x axis) Qa	(y axis)
0.9950	0.7922	1.2254	0.9940	0.7914	0.7665
0.9912	1.0166	1.5819	0.9902	1.0155	0.9896
0.9892	1.1139	1.7329	0.9882	1.1128	1.0840
0.9873	1.2040	1.8718	0.9863	1.2028	1.1709
0.9774	1.5790	2.4507	0.9764	1.5774	1.5331
Qstd slope (m) = 1.55515			Qa slope (m) = 0.97381		
intercept (b) = -0.00218			intercept (b) = -0.00137		
coefficient (r) = 0.99994			coefficient (r) = 0.99994		

y axis =  $\text{SQRT}[\text{H2O}(\text{Pa}/760) (298/\text{Ta})]$

y axis =  $\text{SQRT}[\text{H2O}(\text{Ta}/\text{Pa})]$

CALCULATIONS

Vstd = Diff. Vol [(Pa-Diff. Hg)/760] (298/Ta)  
 Qstd = Vstd/Time

Va = Diff Vol [(Pa-Diff Hg)/Pa]  
 Qa = Va/Time

For subsequent flow rate calculations:

Qstd =  $1/m\{[\text{SQRT}(\text{H2O}(\text{Pa}/760) (298/\text{Ta}))] - b\}$   
 Qa =  $1/m\{[\text{SQRT} \text{H2O}(\text{Ta}/\text{Pa})] - b\}$



# Calibration complies with ISO 17025



Cert. No.:4185-1049515

## Traceable® Certificate of Calibration for Digital Hygrometer/Thermometer

### Instrument Identification:

Model: 11-661-18      S/N: 41531319      Manufacturer : Control Company

### Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Digital Thermometer	41334977/41335007	9/24/05	4000:542181
Chilled Mirror Hygrometer	414	10/22/05	4136 (200562-0)

### Certificate Information:

Technician: 61      Procedure: CAL-17      Cal Date: 1/28/05      Cal Due: 1/28/07  
Test Conditions: 24.0°C      39.0 %RH      1016 mBar

### Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±uc	TUR
°C		N.A.		27.49	27.3	Y	26.5	29.5	0.09	>4:1
%RH		N.A.		20.63	24.5	Y	16.6	24.8	0.88	>4:1
%RH		N.A.		39.10	40.0	Y	37.1	41.1	0.88	2.3:1
%RH		N.A.		72.86	73.8	Y	68.9	76.9	0.88	>4:1

### This Instrument was calibrated using Instruments Traceable to National Institute of Standards and Technology.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 at approximately a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results obtained herein relate only to the item calibrated. This certificate shall not be reproduced except in full.

*Wallace Berry*  
Wallace Berry, Technical Manager

### Maintaining Accuracy:

In our opinion once calibrated your Digital Hygrometer/Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Hygrometer/Thermometers always little, if any at all, but can be affected by aging, temperature, shock, and contamination.

### Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company

**CONTROL COMPANY 4455 Rex Road Friendswood, TX 77546 USA**  
Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com

Control Company is an ISO 17025 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750-01.  
Control Company is ISO 9001 Quality Certified by (DNV) Det Norske Veritas, Certificate No. DPRE-01825-AQ 1103



**R.M. Young Company**  
2801 Acro Park Drive  
Traverse City, Michigan 49686 USA

### Certificate of Calibration and Testing

<b>Test Unit:</b>			
Model:	18802	Serial Number:	CA02612
Description:	Anemometer Drive - 200 to 15,000 Rpm - Comprised of Models 18820A Control Unit & 18830A Motor Assembly		

R.M. Young Company certifies that the above equipment has been inspected and calibrated using standards whose accuracies are traceable to the National Institute of Standards and Technologies (NIST).

Nominal Motor Rpm	27106D Output Frequency Hz (1)	Calculated Rpm (2)	Indicated Rpm (3)
300	50	300	300
2700	450	2700	2700
5100	850	5100	5100
7500	1250	7500	7500
10,200	1700	10200	10200
12,600	2100	12600	12600
15,000	2500	15000	15000

Clockwise and Counterclockwise rotation verified

- (1) Measured frequency output of RM Young Model 27106D standard anemometer attached to motor shaft  
(2) 27106D produces 10 pulses per revolution of the anemometer shaft  
(3) Indicated on the Control Unit LCD display

\*Indicates out of tolerance

No Calibration Adjustments Required       As Found       As Left

Traceable frequency meter used in calibration      DP4863

Date of inspection      13 April 2005

Tested By

## Certificate of Accuracy

### Transfer Standard Type: Barometric Pressure/Altimeter

Transfer standard model: Brunton Multi-Navigator, Version 2.16Serial number: ACP 010796submitted by/owner: Tetra Tech EMI, Boise Idaho

Was compared to Precision Absolute Reference Barometer:

Model number: 355-AJ0900Serial number: 913930-M1Certified accuracy of  $\pm 0.007$ "Hg

NIST traceable to Ruska Deadweight Tester SN 38342/C-85

Date: 4/13/2005

Lab temperature

73.5

°F

Lab pressure

657.6

mm Hg

Reference barometer (in. Hg)	Transfer Standard (in. Hg)	Difference from Reference (in. Hg)	Transfer Standard Correction*
25.89	25.87	-0.02	0.02
26.50	26.49	-0.01	0.01
26.75	26.74	-0.01	0.01
27.00	26.99	-0.01	0.01

**Note:**

If no sign is given on the correction, the true pressure is higher than the indicated pressure. If the sign is negative, the true pressure is lower than the indicated pressure.

Transfer Standard adjustments made? YES  NO 

Post-calibration measurements:

Reference barometer (in. Hg)	Transfer Standard (in. Hg)	Difference from Reference (in. Hg)	Transfer Standard Correction*

Reviewed: RJFDate: 4-13-05

### Chinook Engineering

a division of Inter-Mountain Laboratories, Inc.

555 Absaraka Street

Sheridan, Wyoming 82801 USA

(307) 672-7790

chinook@imlinc.com

1-10-06

1150 Arrive at Met Tower  
1200 Check True North Vane  
Orientation  
1205 Met Tower Offline

GPS Check = 359-10

Sighting Compass = 3450 uncorrected,  
0.6° corrected

Corrected Compass = 0.0°

Barometric Pressure:

Audit = 25061" Hg = 650.49 mmHg

Sensor = 869.2 mb = 652.0 mmHg

Precipitation:

Left Bucket → TIP 1 = 8.6

TIP 2 = 8.4

TIP 3 = 8.3

Rt. Bucket → TIP 1 = 8.0

TIP 2 = 8.0

TIP 3 = 8.1

Wind Speed/Direction:

1. Wind Speed Linearity = Passed (Data on Excel sheet)

2. Wind direction = " " " "

3. Wind speed starting torque = (Ave) 7 g/cm

4. Wind direction starting torque = N/A

1-10-06 (Continued)

Temp: Audit = 53.1

Sensor = 53.6

Humidity: Audit = 40.1%

Sensor = 40.4%

Solar Radiation

Sensor As Found

Zery (Capped) = 0.00 w/m<sup>2</sup>

Reading #1 = 0.434 / 0.865

" #2 = 0.432 / 0.864

" #3 = 0.431 / 0.862

Audit Sensor As Found

Zery Capped = 0.00 w/m<sup>2</sup>

Reading #1 = 0.419

" #2 = 0.418

" #3 = 0.418

1355 Audit Completed → All parameters Passed

1356 Met Tower ONLINE

1415 Leave site

OBSERVATIONS:

1. All sensors functioning properly

2. Wind sensor tail showing wear -

Replacement Recommended

3. w/ Assistance from BDC, slight adjustment

made to Precip. sensor set screw

**Project: Yerington/ARC AQ AM1 PM10 Site**

**Monitor:** Tisch Environmental PM10 Hi-Vol Sampler Model # TE-6070D

Sampler Serial # **613**  
 Orifice Serial # **1013**

Ta Temp C **13.7**  
 Ta Temp K **286.85**  
 Pa Bp mm Hg **649.732**  
 Pa Bp in Hg **25.58**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Date: January 9-10, 2006

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.02-1.24)	Chart Response	IC Corrected	Linear Regression	
1	3.2	1.2220	37	24.6	Slope=	21.07347812
2	3.0	1.1832	36	23.9	Intercept=	-1.1154
3	2.7	1.1226	34	22.6	Corr. Coeff.=	0.9967
4	2.6	1.1016	33	21.9	SFR=	1.2196
5	2.4	1.0584	32	21.3	SSP=	37.0029

**Project: Yerington/ARC AQ AM1 PM10 DUP Site**

**Monitor:** Tisch Environmental PM10 Hi-Vol Sampler Model # TE-6070D

Sampler Serial # **616**  
 Orifice Serial # **1022**

Ta Temp C **12.9**  
 Ta Temp K **286.05**  
 Pa Bp mm Hg **649.732**  
 Pa Bp in Hg **25.58**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Date: January 9-10, 2006

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.02-1.24)	Chart Response	IC Corrected	Linear Regression	
1	2.2	1.0134	31	20.6	Slope=	26.83514709
2	2.4	1.0584	32	21.3	Intercept=	-6.8445
3	2.6	1.1016	34	22.6	Corr. Coeff.=	0.9947
4	2.9	1.1634	37	24.6	SFR=	8.0000
5	3.2	1.2220	39	25.9	SSP=	313.2336

**Project: Yerington/ARC AQ AM1 TSP site**

**Monitor:** Tisch Environmental TSP Hi-Vol Sampler Model # TE-5170D

Sampler Serial # **1033**

Ta Temp C **9.8**  
 Ta Temp K **282.95**  
 Pa BP mm Hg **649.732**  
 Pa Bp in Hg **25.58**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Date: January 9-10, 2006

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.10-1.70)	Chart Response	IC Corrected	Linear Regression	
1	4	1.3660	39	25.9	Slope=	18.59464845
2	4.2	1.3997	40	26.6	Intercept=	0.5018
3	4.7	1.4806	42	27.9	Corr. Coeff.=	0.9992
4	5.1	1.5423	44	29.2	SFR=	1.2031
5	5.6	1.6161	46	30.6	SSP=	34.6592

**Project: Yerington/ARC AQ AM2 PM10 site**

**Monitor:** Tisch Environmental PM10 Hi-Vol Sampler Model # TE-6070D

Sampler Serial # **617**  
 Orifice Serial # **1016**

Ta Temp C **9.5**  
 Ta Temp K **282.65**  
 Pa Bp mm Hg **649.732**  
 Pa Bp in Hg **25.58**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Date: January 9-10, 2006

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.02-1.24)	Chart Response	IC Corrected	Linear Regression	
					Slope=	Intercept=
1	3.1	1.1939	36	23.7	Slope=	23.14822554
2	3.0	1.1745	35	23.1	Intercept=	-4.0203
3	2.6	1.0935	32	21.1	Corr. Coeff.=	<b>0.9966</b>
4	2.4	1.0507	31	20.4	SFR=	1.2018
5	2.3	1.0286	30	19.8	SSP=	36.0824

**Project: Yerington/ARC AQ AM2 TSP site**

**Monitor:** Tisch Environmental TSP Hi-Vol Sampler Model # TE-5170D

Sampler Serial # **NA**  
 Orifice Serial # **1034**

Ta Temp C **9.5**  
 Ta Temp K **282.65**  
 Pa BP mm Hg **649.732**  
 Pa Bp in Hg **25.58**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Date: January 9-10, 2006

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.10-1.70)	Chart Response	IC Corrected	Linear Regression	
					Slope=	Intercept=
1	3.7	1.3042	38	25.1	Slope=	23.25564999
2	4.3	1.4059	41	27.0	Intercept=	-5.3475
3	4.9	1.5007	45	29.7	Corr. Coeff.=	<b>0.9941</b>
4	5.2	1.5459	47	31.0	SFR=	1.2018
5	5.6	1.6042	48	31.7	SSP=	34.2658

**Project: Yerington/ARC AQ AM3 PM10 site**

**Monitor:** Tisch Environmental PM10 Hi-Vol Sampler Model # TE-6070D

Sampler Serial # **618**  
 Orifice Serial # **1018**

Ta Temp C **9.1**  
 Ta Temp K **282.25**  
 Pa Bp mm Hg **653.034**  
 Pa Bp in Hg **25.71**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Date: January 9-10, 2006

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.02-1.24)	Chart Response	IC Corrected	Linear Regression	
1	3.3	1.2185	39	25.6	Slope=	20.0859134
2	3.0	1.1707	37	24.3	Intercept=	1.0636
3	2.7	1.1107	36	23.7	Corr. Coeff.=	<b>0.9901</b>
4	2.5	1.0689	34	22.4	SFR=	1.1940
5	2.3	1.0253	33	21.7	SSP=	38.0974

**Project: Yerington/ARC AQ AM3 TSP site**

**Monitor:** Tisch Environmental TSP Hi-Vol Sampler Model # TE-5170D

Sampler Serial # **NA**  
 Orifice Serial # **1014**

Ta Temp C **9.4**  
 Ta Temp K **282.55**  
 Pa BP mm Hg **653.034**  
 Pa Bp in Hg **25.71**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Date: January 9-10, 2006

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.10-1.70)	Chart Response	IC Corrected	Linear Regression	
1	2.8	1.1311	35	23.0	Slope=	19.72410619
2	3.1	1.1901	37	24.3	Intercept=	0.7259
3	3.7	1.3000	40	26.3	Corr. Coeff.=	<b>0.9995</b>
4	4.1	1.3684	42	27.6	SFR=	1.1953
5	5.1	1.5260	47	30.9	SSP=	36.9450

**Project: Yerington/ARC AQ AM4 PM10 site**

**Monitor:** Tisch Environmental PM10 Hi-Vol Sampler Model # TE-6070D

Sampler Serial # **619**  
 Orifice Serial # **1012**

Ta **3.1**  
 Ta **276.25**  
 Pa **653.034**  
 Pa **25.71**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Date: January 9-10, 2006

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.02-1.24)	Chart Response	IC Corrected	Linear Regression	
1	3.3	1.2147	39	25.4	Slope=	23.08031607
2	3.1	1.1774	37	24.1	Intercept=	-2.8935
3	2.8	1.1190	35	22.8	Corr. Coeff.=	<b>0.9943</b>
4	2.6	1.0784	34	22.1	SFR=	1.1686
5	2.4	1.0253	32	20.8	SSP=	37.0211

**Project: Yerington/ARC AQ AM4 TSP site**

**Monitor:** Tisch Environmental TSP Hi-Vol Sampler Model # TE-5170D

Sampler Serial # **NA**  
 Orifice Serial # **1015**

Ta **3.1**  
 Ta **276.25**  
 Pa **653.034**  
 Pa **25.71**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Date: January 9-10, 2006

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.10-1.70)	Chart Response	IC Corrected	Linear Regression	
1	2.9	1.1388	35	22.8	Slope=	21.46164197
2	3.5	1.2509	39	25.4	Intercept=	-1.6413
3	4.1	1.3538	42	27.3	Corr. Coeff.=	<b>0.9990</b>
4	4.9	1.4799	46	29.9	SFR=	1.1686
5	5.4	1.5535	49	31.9	SSP=	36.0380

**Project: Yerington/ARC AQ AM5 PM10 site**

**Monitor:** Tisch Environmental PM10 Hi-Vol Sampler Model # TE-6070D

Sampler Serial # **614**  
 Orifice Serial # **1019**

Ta Temp C **7.7**  
 Ta Temp K **280.85**  
 Pa Bp mm Hg **653.034**  
 Pa Bp in Hg **25.71**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Date: January 9-10, 2006

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.02-1.24)	Chart Response	IC Corrected	Linear Regression	
					Slope=	Intercept=
1	2.3	1.0227	32	21.0	20.20337557	
2	2.6	1.0873	34	22.3	0.3414	
3	2.8	1.1182	35	23.0	<b>Corr. Coeff.= 0.9993</b>	
4	2.9	1.1482	36	23.6	SFR=	1.3279
5	3.1	1.1871	37	24.3	SSP=	41.4284

**Project: Yerington/ARC AQ AM5 TSP site**

**Monitor:** Tisch Environmental TSP Hi-Vol Sampler Model # TE-5170D

Sampler Serial # **NA**  
 Orifice Serial # **1017**

Ta Temp C **7.7**  
 Ta Temp K **280.85**  
 Pa BP mm Hg **653.034**  
 Pa Bp in Hg **25.71**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Date: January 9-10, 2006

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.10-1.70)	Chart Response	IC Corrected	Linear Regression	
					Slope=	Intercept=
1	3.1	1.1871	38	24.9	19.50137328	
2	3.5	1.2613	41	26.9	2.0185	
3	3.9	1.3313	42.5	27.9	<b>Corr. Coeff.= 0.9962</b>	
4	4.3	1.3979	45	29.5	SFR=	1.1881
5	5.3	1.5518	49	32.1	SSP=	38.4078

**Project: Yerington/ARC AQ AM6 PM10 site**

**Monitor:** Tisch Environmental PM10 Hi-Vol Sampler Model # TE-6070D

Sampler Serial # **615**  
 Orifice Serial # **1020**

Date: January 9-10, 2006

Ta Temp C **9.8**  
 Ta Temp K **282.95**  
 Pa Bp mm Hg **654.812**  
 Pa Bp in Hg **25.78**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.02-1.24)	Chart Response	IC Corrected	Linear Regression	
1	2.4	1.0472	34	22.3	Slope=	23.19629628
2	2.6	1.0899	35	23.0	Intercept=	-2.0529
3	2.8	1.1309	37	24.3	Corr. Coeff.=	<b>0.9930</b>
4	3.0	1.1706	38	25.0	SFR=	1.1937
5	3.1	1.1899	39	25.6	SSP=	39.0004

**Project: Yerington/ARC AQ AM6 TSP site**

**Monitor:** Tisch Environmental TSP Hi-Vol Sampler Model # TE-5170D

Sampler Serial # **NA**  
 Orifice Serial # **1021**

Date: January 9-10, 2006

Ta Temp C **9.8**  
 Ta Temp K **282.95**  
 Pa BP mm Hg **654.812**  
 Pa Bp in Hg **25.78**

Orifice Serial # **w43**  
 W43 slope (m) **0.97381**  
 W43 int. (b) **-0.00137**

Point	Orifice Press. Drop (" H2O)	Qa (Orifice) (1.10-1.70)	Chart Response	IC Corrected	Linear Regression	
1	3.5	1.2643	38	25.0	Slope=	19.89048839
2	3.9	1.3345	40	26.3	Intercept=	-0.1468
3	4.2	1.3848	42	27.6	Corr. Coeff.=	<b>0.9979</b>
4	4.7	1.4648	44	28.9	SFR=	1.1937
5	4.9	1.4956	45	29.6	SSP=	35.8968